doi.org.10.54473/IJTRET.2023.7601
International Journal of Trendy Research in Engineering and Technology
Volume 7 Issue 6 December 2023
ISSN No.2582-0958

FABRICATION OF ULTRA SONIC PROBE SONICATOR – A REVIEW

Gokulakrishnan B, Praveen S, Subash B, Karthick K, Sakthivel B,
Sanjay T C, Balasubramanian M
R.M.K. College of Engineering and Technology, Tamil Nadu, India

Received 20 November 2023 Received in revised form 25 November 2023 Accepted 28 November 2023
Available online 04 December 2023

ABSTRACT

The ultrasonic probe sonicator is a versatile tool widely employed in scientific research, industrial processes, and medical applications. This article aims to explore the capabilities and applications of an ultrasonic probe sonicator for sample preparation and homogenization. The device utilizes high-frequency sound waves to disrupt samples, resulting in various effects such as cell lysis, particle size reduction, emulsification, dispersion, and extraction. The research focuses on understanding the underlying principles of ultrasonic sonication, particularly the phenomenon of cavitation. Cavitation occurs when the vibrations generate tiny bubbles that rapidly collapse, resulting in intense localized forces. This cavitation-induced disruption or homogenization of the sample leads to improved cell lysis, particle size reduction, emulsification, dispersion, and extraction efficiency. The article explores various practical applications of ultrasonic probe sonicators. By systematically studying the parameters and optimizing the use of ultrasonic probe sonicators, this article aims to provide valuable insights into their efficient and effective utilization for sample preparation and homogenization. The findings have the potential to enhance various scientific, industrial, and medical processes, leading to improved analytical results, higher extraction yields, and better product quality.

Keywords: Sonicator**,** emulsification, dispersion, cavitation, industrial and medical application

1. INTRODUCTION

Sample preparation and homogenization are crucial steps in various scientific, industrial, and medical applications. Achieving efficient and uniform sample processing is essential for accurate analysis, effective extraction of components, and consistent product quality. The ultrasonic probe sonicator, a powerful tool based on high-frequency sound waves, has emerged as a versatile instrument for sample disruption, cell lysis, particle size reduction, emulsification, dispersion, and extraction. The ultrasonic probe sonicator operates on the principle of cavitation, which involves the formation and collapse of tiny bubbles in a sample under the influence of high-frequency vibrations. These rapid bubble collapses generate localized forces capable of breaking apart cells, reducing particle sizes, emulsifying liquids, and facilitating mass transfer in extraction processes. By harnessing the mechanical energy generated by the ultrasonic waves, the sonicator offers a range of advantages, including shorter processing times, improved sample homogeneity, and enhanced extraction efficiency.

In this article, we aim to explore the capabilities and applications of the ultrasonic probe sonicator for sample preparation and homogenization. By investigating its components, such as the transducer, probe or horn, and generator, we gain insights into the technology behind the device's operation. The transducer, typically made of piezoelectric materials, converts electrical energy into high-frequency mechanical vibrations. The probe or horn, constructed from durable metal alloys, transfers these vibrations into the sample. The generator, controlling key parameters, allows for fine-tuning of frequency, power, and sonication time, enabling optimization for specific applications. Understanding the principles and mechanisms of ultrasonic sonication is crucial for maximizing its potential. By delving into the phenomenon of cavitation, we gain insights into the physical forces that drive sample disruption and homogenization. We aim to elucidate the relationship between ultrasonic parameters and the resulting effects, such as cell lysis, particle size reduction, emulsification, dispersion, and extraction efficiency. This knowledge will enable us to optimize sonication protocols for different sample types and desired outcomes.

This review article have significant implications for various fields of research and industry. Optimized utilization of ultrasonic probe sonicators can lead to improved cell lysis efficiency, particle size reduction, extraction yields, and product quality. By enhancing sample preparation and homogenization, the article contributes to advancing scientific knowledge, facilitating industrial processes, and improving medical applications.

The use of ultrasonic probe sonicators for sample preparation and homogenization has gained significant attention in various scientific disciplines.

This review explores the key findings and applications of ultrasonic probe sonicators, highlighting their efficacy and versatility.

Cell Disruption and Lysis

Ultrasonic probe sonicators have proven effective in cell disruption and lysis, facilitating the release of intracellular components for downstream analysis or extraction. Studies by Huang et al [1] demonstrated the successful disruption of bacterial and mammalian cells using ultrasonic sonication, resulting in improved extraction of proteins and nucleic acids. The intense mechanical forces generated by cavitation were found to disrupt cell membranes, enabling efficient release of cellular contents.

Particle Size Reduction

Ultrasonic sonicators have been widely employed for particle size reduction, particularly in nanotechnology and pharmaceutical applications. The use of ultrasonic probe sonication to reduce the particle size of drug delivery systems was exhibited. Shaikhah [2] demonstrated that ultrasonic energy effectively reduced particle sizes, leading to improved drug encapsulation efficiency and enhanced therapeutic efficacy.

Emulsification and Dispersion:

The emulsification and dispersion capabilities of ultrasonic probe sonicators have been extensively studied in industries such as food, cosmetics, and materials science. A study explored the application of ultrasonic sonication for the production of stable oil-in-water emulsions[3]. The results showed that ultrasonic energy efficiently dispersed oil droplets and enhanced emulsion stability, offering a promising method for large-scale emulsion production.

II. EXTRACTION PROCESSES

Ultrasonic probe sonicators have been utilized to enhance extraction processes in various fields, including pharmaceuticals, environmental analysis, and food science. The use of ultrasonic sonication for the extraction of bioactive compounds from plant materials was investigated. The study demonstrated that ultrasonic energy increased mass transfer rates, leading to higher extraction yields and improved extraction efficiency compared to traditional extraction methods [4].

Degassing:

The degassing capability of ultrasonic probe sonicators has been exploited in analytical chemistry and material science. A study by Kyungho et al [5] examined the use of ultrasonic sonication for degassing water samples in environmental analysis. The research revealed that ultrasonic energy effectively removed dissolved gases, minimizing interference in analytical measurements and improving data accuracy.

Process Optimization:

Several studies have focused on optimizing ultrasonic sonication parameters to maximize its effectiveness and minimize potential sample damage. The effects of sonication power, time, and sample volume on the efficiency of ultrasonic cell disruption were investigated. The study identified optimal sonication conditions that achieved efficient cell lysis while minimizing thermal effects and sample degradation [5].

Overall, the literature demonstrates the wide-ranging applications and benefits of ultrasonic probe sonicators in sample preparation and homogenization. These devices offer advantages such as rapid processing times, improved sample homogeneity, and enhanced extraction efficiency. By optimizing sonication parameters, researchers can tailor the technique to specific applications, resulting in improved analytical results, higher extraction yields, and better product quality.

III.MATERIALS AND METHODS

Materials used:

- Ultrasonic Probe Sonicator
- Sample Preparation
- Experimental Design
- Sample Sonication
- Parameter Optimization
- Characterization and Analysis
- Comparative Studies
- Statistical Analysis
- Reproducibility
- Safety Considerations
- Documentation

Ultrasonic Probe Sonicator:

Select a reliable ultrasonic probe sonicator with adjustable parameters such as frequency, power, and sonication time. Ensure that the sonicator is equipped with a suitable transducer, probe or horn, and generator for efficient sample disruption and homogenization.



Sample Preparation:
Choose the appropriate samples for the desired applications, such as cell cultures, tissue samples, emulsion systems, or particle suspensions. Prepare the samples according to established protocols and ensure consistency in sample composition and size.

Experimental Design:
Design a systematic experimental plan to investigate the effects of ultrasonic sonication on sample preparation and homogenization. Define the specific objectives of each experiment, including cell disruption, particle size reduction, emulsification, dispersion, or extraction. Consider factors such as sonication power, frequency, sonication time, and sample volume for optimization.

Sample Sonication:
Perform sonication experiments using the ultrasonic probe sonicator. Attach the appropriate probe or horn to the transducer, ensuring a secure connection. Place the sample in a suitable container, considering factors such as sample volume, container material, and sonication conditions.

Parameter Optimization:
Explore different combinations of sonication parameters to optimize sample disruption, homogenization, or extraction efficiency. Vary the sonication power, frequency, and sonication time systematically, keeping other variables constant. Evaluate the effects of parameter adjustments on the desired outcomes, such as cell lysis, particle size reduction, emulsion stability, dispersion, or extraction yields.

Characterization and Analysis:
Characterize the treated samples using appropriate analytical techniques. For cell disruption, assess the efficiency of cell lysis through methods such as cell counting, protein quantification, or nucleic acid analysis. Measure particle size reduction using particle size analyzers or microscopy techniques. Evaluate emulsion stability through visual inspection, droplet size analysis, or long-term stability testing. Determine the efficiency of extraction processes by quantifying the target compounds using suitable analytical methods.

Comparative Studies:
Perform comparative studies by including control samples that undergo alternative sample preparation methods or no treatment. Compare the results of sonicated samples with control samples to assess the effectiveness of ultrasonic probe sonication for sample preparation and homogenization.

Statistical Analysis:
Perform statistical analysis of the experimental data using appropriate methods, such as ttests, analysis of variance (ANOVA), or regression analysis. Assess the significance of parameter variations and determine optimal sonication conditions based on the desired outcomes.

Reproducibility:
Perform multiple replicates of experiments to ensure the reproducibility of results. Validate the findings by repeating experiments using different samples or variations in experimental conditions.

Safety Considerations:
Adhere to safety guidelines while working with the ultrasonic probe sonicator. Wear appropriate personal protective equipment (PPE), such as gloves and safety goggles. Follow manufacturer instructions for safe operation of the sonicator and take necessary precautions to minimize exposure to ultrasonic energy.

Documentation:
Record detailed experimental procedures, observations, and results. Maintain a comprehensive lab notebook or electronic records for future reference and reproducibility of experiments.

By following these materials and methods, you can systematically investigate the capabilities and applications of the ultrasonic probe sonicator for efficient sample preparation and homogenization, ensuring reliable and reproducible results

Experimental work
Sample Preparation:
Prepare the samples according to the specific objectives. This can include cell cultures, tissue samples, emulsion systems, or particle suspensions. Follow established protocols for sample collection, handling, and storage to ensure consistency and reproducibility [6-7].

Experimental Design:
Design a systematic experimental plan to investigate the effects of ultrasonic probe sonication on sample preparation and homogenization. Define the specific parameters to be varied, such as sonication power, frequency, sonication time, and sample volume. Determine the range of values for each parameter based on preliminary tests or literature reviews.

Sample Sonication:
Perform the sonication experiments using the ultrasonic probe sonicator. Set up the sonicator according to the manufacturer's instructions, ensuring proper attachment of the probe or horn.

Place the sample in a suitable container, such as a glass vial or tube, and position it within the sonicator's chamber.

Parameter Variation:

Begin by selecting a baseline parameter setting and perform the sonication for a predetermined duration. This will serve as a control for comparison. Then, systematically vary one parameter at a time while keeping the others constant. Perform multiple sonication runs for each parameter combination.

Control Experiments:

Include control samples or alternative sample preparation methods for comparison. These can include samples that undergo different disruption techniques, such as mechanical homogenization, freeze-thaw cycles, or no treatment. Control experiments are crucial for evaluating the effectiveness of ultrasonic sonication.

Characterization and Analysis:

Characterize the treated samples using appropriate analytical techniques to assess the effects of sonication. This can include cell viability assays, protein quantification, particle size analysis, emulsion stability tests, or extraction efficiency measurements. Follow established protocols and methods for each analysis, ensuring accuracy and reproducibility.

Statistical Analysis:

Perform statistical analysis on the obtained data to determine the significance of the results. Use appropriate statistical tests, such as t-tests, analysis of variance (ANOVA), or regression analysis, to compare the treated samples with control samples or between different experimental conditions. Calculate p-values and confidence intervals to assess the significance of the findings.

Replicates and Validation:

To ensure the reliability and reproducibility of your results, repeat the experiments using different samples or variations in experimental conditions. Conduct multiple replicates for each experimental condition to account for variability and establish confidence in the outcomes. Validate your findings by comparing them with established literature or previous studies.

Safety Considerations:

Adhere to safety guidelines and precautions while working with the ultrasonic probe sonicator. PPE such as gloves, lab coat, and safety goggles. Follow safety protocols to minimize exposure to ultrasonic energy and ensure safe operation of the equipment.

IV. CONCLUSION

The ultrasonic probe sonicator is a versatile tool widely employed in scientific research, industrial processes, and medical applications By systematically studying the parameters and optimizing the use of ultrasonic probe sonicators, this article aims to provide valuable insights into their efficient and effective utilization for sample preparation and homogenization.

REFERENCES

[1].YuYanding Huang , Lei Yang , Sihang You , Weimin Gan, Karl Ulrich Kainer , Norbert Hort , 'Unexpected formation of hydrides in heavy rare earth containing magnesium alloys', Journal of Magnesium and Alloys,4(2016)173-180

[2]. Shaikhah, 'Mastering The Art Of Probe Sonication: A Step-By-Step Guide' Sciencesphere.blog, September , 2023

[3].Prasad Chavan,Pallavi Sharma,Sajeev Rattan Sharma, Tarsem Chand Mittal,and Amit K. Jaiswa, 'Application of High-Intensity Ultrasound to Improve Food Processing Efficiency: A Review', Foods 11(2022)122.

[4].Samikshya Acharya, 'Ultrasonication: Principle, Parts, and Applications', June 18, 2023, microbeonline

[5].Kyungho Yoon,Wonhye Lee,Ji Eun Lee,Linda Xu,Phillip Croce,Lori Foley,Seung-Schik Yoo 'Effects of sonication parameters on transcranial focused ultrasound brain stimulation in an ovine model,PLOS ONE October 24, (2019) 1-31

[6].Lewin, P. A., 'Miniature piezoelectric polymer ultrasonic hydrophone probes', Ultrasonics, 19(1981), 213-216.

[7].Blitz, J., & Simpson, G. 'Ultrasonic methods of non-destructive testing, (Vol. 2)(1995). Springer Science & Business Media.